GOOD WILL INSTRUMENT CO., LTD. N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

G<u><u></u></u>INSTEK

GPT-12000 Specifications

The specifications apply when the GPT-12000 is powered on for at least 30 minutes under +15°C~+35°C

GPT-12004 (Front)







GPT-12004 (Rear)



GPT-12003/12002/12001 (Rear)



(All models are available with optional GPIB or LAN)

Model \ Func.	AC Withstanding	DC Withstanding	Insulation Resistance	Ground Bond	Ground Continuity
GPT-12001	v				V
GPT-12002	v	v			V
GPT-12003	v	V	V		V
GPT-12004	V	V	v	V	V

AC WITHSTANDING	
Output-Voltage Range	0.050kV~5.000kV
Output-Voltage Resolution	1V
Output-Voltage Accuracy	\pm (1% of setting + 5V) [no load]
Maximum Rated Load	200 VA (5kV/40mA)
Maximum Rated Current	40mA (0.5kV< V ≦5kV)
	10mA (0.05kV \leq V \leq 0.5kV)
Output-Voltage Waveform	Sine wave
Output-Voltage Frequency	50 Hz / 60 Hz selectable
Voltage Regulation	\pm (1% + 5V) [maximum rated load $ ightarrow$ no load]
Voltmeter Accuracy	\pm (1% of reading + 5V)
Current Measurement Range	1μA~40.00mA
Current Best Resolution	1µА / 10µА
Current Measurement Accuracy	\pm (1.5% of reading + 30 μ A)
Current Offset	60μA Maximum
Window Comparator Method	Yes
ARC Detect	Yes
RAMP UP (Rise Time)	0.1s~999.9s
RAMP DOWN (Fall Time)	0.0s~999.9s

GPT12000_Spec_E1.doc

GOOD WILL INSTRUMENT CO., LTD.

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

G<u><u></u></u>INSTEK

TIMER (Test Time)*		OFF, 0.3s~999.9s	
TIMER Accuracy		±(100ppm + 20ms)	
WAIT TIME		0.0s~999.9s	
GND		ON/OFF	
DC WITHSTANDING			
Output-Voltage Range		0.050kV~6.000kV	
Output-Voltage Resolution		1V	
Output-Voltage Accuracy		± (1% of setting + 5V) [no load]	
Maximum Rated Load		50W (5kV/10mA)	
Maximum Rated Current	t	10mA (0.5kV< V ≦6kV)	
		$2mA(0.05kV \le V \le 0.5kV)$	
Voltage Regulation		\pm (1% + 5V) [maximum rated load \rightarrow no load]	
Voltage Regulation Voltmeter Accuracy		\pm (1% of reading + 5V)	
Current Measurement Range		1μΑ~10.00mA	
Current Best Resolution		0.1μΑ /1μΑ /10μΑ	
	couracy	\pm (1.5% of reading + 3µA) when I Reading < 1mA	
Current Measurement Accuracy		\pm (1.5% of reading + 30µA) when I Reading \geq 1mA	
Current Offset		$\pm (1.5\% \text{ of reading} \pm 50\mu\text{A})$ when reading $\geq 111\text{A}$ 5 μA Maximum	
	athod		
Window Comparator Me	ethou	Yes	
ARC Detect		Yes	
RAMP UP (Rise Time)	1	0.1s~999.9s	
RAMP DOWN (Fall Time)	0.0s~999.9s	
TIMER (Test Time)*		OFF, 0.3s~999.9s	
TIMER Accuracy		±(100ppm + 20ms)	
WAIT TIME		0.0s~999.9s	
GND	_	ON/OFF	
INSULATION RESISTANCE			
Output Voltage		50V~1200V dc	
Output-Voltage Resolution		50V	
Output-Voltage Accuracy		± (1% of setting + 5V) [no load]	
Resistance Measuremen			
Test Voltage	Display Range	Measurement Range / Accuracy	
50V≦V≦100V	0.1MΩ~ 10.00GΩ	$0.1M\Omega^{-1}M\Omega$: ±(5% of reading + 3 count)	
150V≤V≤450V	0.1MΩ~ 20.00GΩ	1 MΩ~50MΩ : \pm (5% of reading + 1 count)	
15002024500		51MΩ~2GΩ:±(10% of reading + 1 count)	
500V≦V≦1200V	0.1MΩ~ 50.00GΩ	$0.1M\Omega^{\sim}1M\Omega$: ±(5% of reading + 3 count)	
		1 MΩ~500MΩ \div ±(5% of reading + 1 count)	
		501MΩ~9.999GΩ:±(10% of reading + 1 count)	
		$10G\Omega^{5}0G\Omega$: ±(20% of reading + 1 count)*	
Voltage Regulation		\pm (1% + 5V) [maximum rated load \rightarrow no load]	
Voltmeter Accuracy		\pm (1% of reading + 5V)	
Short-Circuit Current		10mA max.	
Output Impedance		2kΩ	
Window Comparator Method		Yes	
RAMP UP (Rise Time)		0.1s~999.9s	
RAMP DOWN (Fall Time)			
		0.0s~999.9s OFF, 0.3s~999.9s	
TIMER (Test Time)*			
		±(100ppm + 20ms)	
WAIT TIME GND		0.0s~999.9s ON/OFF	
-			
Ground Bond		02.004~22.004.20	
Output-Current		03.00A~32.00A ac	
Output-Current Resolution		0.01A	
Output-Current Accuracy		$3A \le I \le 8A : \pm (1\% \text{ of reading} + 0.2A)$	
		8A <i≦ +="" 0.05a)<="" 32a="" :="" of="" reading="" td="" ±(1%=""></i≦>	
Test-Voltage		8Vac max (open circuit)	
GPT12000 Spec F1 doc		2 OF 3	

GPT12000_Spec_E1.doc

GOOD WILL INSTRUMENT CO., LTD.

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

Test-Voltage Frequency	50Hz/60Hz selectable	
Ohmmeter Measurement Range	$1 \text{m}\Omega^{\sim}$ 650m Ω	
Ohmmeter Measurement Resolution	0.1mΩ	
Ohmmeter Measurement Accuracy	$\pm (1\% \text{ of reading} + 2 \text{ m}\Omega)$	
Window Comparator Method	Yes	
TIMER (Test Time)	0.3s~999.9s	
TIMER Accuracy	±(100ppm + 20ms)	
Test Method	Four Terminal	
GND	ON/OFF	
Continuity Test		
Output-Current	100mA dc (fixed)	
Ohmmeter Measurement Range	0.10Ω~ 70.00Ω	
Ohmmeter Measurement Resolution	0.01Ω	
Ohmmeter Measurement Accuracy	$\pm(10\% \text{ of reading} + 2 \Omega)$	
Window Comparator Method	Yes	
TIMER (Test Time)*	0.3s~999.9s	
TIMER Accuracy	±(100ppm + 20ms)	
MEMORY		
Single Step Memory	MANU: 100 blocks	
Automatic Testing Memory	AUTO : 100 blocks, manu per auto : 10	
INTERFACE		
REMOTE (Front) terminal	Standard	
USB host (Front)	Standard	
Rear Output	Standard	
RS-232C	Standard	
USB device	Standard	
Signal I/O	Standard	
GPIB	Option	
LAN	Option	
DISPLAY		
	7" color LCD	
POWER SOURCE		
	AC 100V~240V ± 10%, 50Hz/60Hz	
DIMENSION & WEIGHT		
GPT-12001/12002/12003	380(W) x 148(H) x 436(D) mm; Approx. 11kg	
GPT-12004	380(W) x 148(H) x 454(D) mm; Approx. 15kg	

* When Ground Mode is "ON", the measurement range is 30GΩ max. and adding 10% error for accuracy.