

# Certificate of Compliance

**Certificate:** 1188789

**Master Contract:** 204767

**Project:** 1188789

**Date Issued:** August 17, 2001


**Issued to:** **Camille Bauer AG**  
Aargauerstrasse 7  
CH-5610 Wohlen  
SWITZERLAND

**Attention:** Mr. Volker Pohl

*The products listed below are eligible to bear the CSA Mark shown  
with adjacent indicators 'C' and 'US'*



**Issued by:**   
Enzo De Simone, B.A.Sc.

**Authorized by:**   
Nick Alfano  
Operations Manager

## PRODUCTS

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - CERTIFIED TO U.S. STANDARDS

Programmable Multi-Transducer, Model SINEAX M 563, fixed equipment, rated 24-60Vdc/ac, 50/60Hz, 7VA or 85-230Vdc/ac, 50/60Hz, 7VA, or powered from measuring inputs, Equipment Class II, Pollution Degree 2, Installation Category III.

## APPLICABLE REQUIREMENTS

- |                              |   |   |
|------------------------------|---|---|
| CAN/CSA-C22.2 No. 1010.1-92  | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)                        |
| CAN/CSA-C22.2 No. 1010.1B-97 | - | Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements" |
| UL Std No. 3111-1            | - | Equipment for Measurement Use; Part I: General Requirements   |

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S. respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.

## *Supplement to Certificate of Compliance*

**Certificate:** 1188789

**Master Contract:** 204767

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1188789	2001-08-17	Original Certification.