

EC Declaration of Conformity

Pico Technology declares that the following products comply with the requirements of the specified Directives and Standards as listed below. Technical documentation required to demonstrate compliance to the standards is available for inspection by the relevant enforcement authorities. Products carry the CE mark.

Products covered by this declaration:

PicoScope 9301 2 channel USB sampling oscilloscope.
PicoScope 9302 2 channel USB sampling oscilloscope with clock recovery.
PicoScope 9311 2 channel USB sampling scope with differential 60ps TDR.
PicoScope 9312 2 channel USB sampling scope with differential 40ps TDR.
PicoScope 9321 2 channel USB sampling oscilloscope with optical to electrical converter and clock recovery.
PicoScope 9341 4 channel USB sampling oscilloscope.

EU Directives covered by this declaration:

2004/108/EC - Electromagnetic Compatibility Directive.
2006/95/EC - Low Voltage Equipment Directive.

The basis on which conformity is being declared:

EN61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use, general equipment requirements.

EN61010-2-030:2010 Particular requirements for testing and measuring circuits.

EN61326-1:2006 Electrical equipment for measurement, control and laboratory use – EMC requirements. Group 1, Class A equipment – (emissions section only)

EN61326-1:2006 Electrical equipment for measurement, control and laboratory use – EMC requirements. Basic Immunity – (immunity section only)

CFR 47:2009 Code of Federal Regulations FCC: part 15 Subpart B – Frequency devices – unintentional Radiators. Radiated emissions standard. Class A emissions.

EN61000-3-3:2008 Electromagnetic compatibility (EMC) - Part 3-3: limits – Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current <16A per phase and not subject to conditional connection.

EN61000-3-2:2006 Electromagnetic compatibility (EMC) – Part 3-2 Limits – Limits for harmonic current emissions for equipment input current up to and including 16A per phase.

Alan Tong
Managing Director
03 Nov 2014

Signed

